

MINUTES OF PRE BID MEETING

Held on 19.05.2017

**Supply, Installation, Commissioning & Maintenance of Food Technology
(Nutrigenomics) Laboratory Equipment at IKGPTU Kapurthala and its other constituent
campuses (Punjab)**

NIT No: EdCIL/PROC/PTU-2017/LAB-MTECH-FOODTECH/FOODTECH-PKG1/1



EdCIL (India) Limited

A Mini Ratna Company

[A Government of India Enterprise]

Under Ministry of Human Resource Development

Plot 18 A, Sector 16 A, NOIDA – 201 301

Uttar Pradesh, INDIA

Tel: 91-120-2512001 to 2512006 (EPABX)

Tel: 91-120-4310840 (Direct)

Fax: 91-120-2515372, 2515485

Note: This document is serially numbered from pages 01 to 5.

Pre Bid Meeting Date: 19.05.2017

Queries Regarding Commercial Specifications:

S. No.	Clause No./ Annexure/ Item	Page No.	Name of Company / Agency who had raised queries	Bidder's Query	Clarification
1.	EMD	Page 5 (Technical Bid Document)	M/s.S V Instruments Analytica Pvt. Ltd	EMD you are asking for Bank Drat only, we request you to make it either bank draft or Bank Guarantee. All govt./PSUs prefer for bank guarantee. And more over validity of bank draft is 90 days only while bank guarantee has more validity.	"NO Change in Tender Condition"
2.	Duties (Custom & Excise)	Page 28 (Technical Bid Document)	M/s.S V Instruments Analytica Pvt. Ltd	If we do quote all equipments (as you said in your tender document, one has to quote one and all the equipment) as a package in Indian Rupees as a importer, how do we get Custom Exemption and if there is any format for getting Custom Duty Exemption while quoting in INR.	IKGPTU is exempted from paying custom duty and necessary certificate may be issued as per the requirement.
3.	Price Schedule Form 2A & Form 2B	Page 4 & 5 (Financial Bid Document)	M/s.S V Instruments Analytica Pvt. Ltd	Verbal Query received regarding change of format of price schedule in Form 2A: "Change F.O.B and C.I.F terms to C.I.F and C.I.P Terms respectively for goods to reach destination without transferring the risk to buyer."	Price Schedule Form 2A & Form 2B with Suggested modification is attached as Corrigendum No:- 02.
			M/s Eppendorf	Verbal Query received regarding change of format of price schedule in Form 2B: " Change C.I.F term to F.O.R term for goods to reach destination without transferring the risk to buyer."	

Queries Regarding Technical Specifications:

S.No	Page No.	Items		Existing Specification	Name of Company / Agency who had raised queries	Bidder's Query	Clarifications
1.	37 (Annexure I of Technical BID) 8 (Annexure IV of Financial Bid)	3.Refrigerated Centrifuge	8. Power Consumption	Max.500 W	M/s Eppendorf	Should change to 1100 W or better	“1000 W or better “
2.	& 16 (Technical Compliance Report of Financial Bid)		9. Dimension (W x D x H)	300 x 460 x250 mm	M/s Eppendorf	Should change to 380 x 640 x 290 mm or better	“Dimension Point is removed”
3.	37 (Annexure I of Technical BID) 8 (Annexure IV of Financial Bid)	4.Centrifuge	2. 8 x 15 ml rotor Head and 4x 50 ml rotor head or better	2. 8 x 15 ml rotor Head and 4x 50 ml rotor head or better	M/s Eppendorf	Should change to 6 x15 ml rotor head and 4 x 50 ml rotor head or better	“6 x15 ml rotor head and 4 x 50 ml rotor head or better”

	& 17 (Technical Compliance Report of Financial Bid)						
4.	43 (Annexure I of Technical BID) 12-13 (Annexure IV of Financial Bid)	13. CO ₂ Incubator	4. Jacket & Sensor	Air Jacketed Heating system with T/c Sensor	M/s Eppendorf	Should be "Air Jacketed Heating system with TC/IR Sensor"	Air Jacketed Heating system with TC/IR Sensor
5.	& 22 (Technical Compliance Report of Financial Bid)		6. O ₂ - Concentration	1 to 20% or better	M/s Eppendorf	Should be 1 to 19% or better	"1 to 19 % or better"
6.	41 (Annexure I of Technical BID) 11 (Annexure IV of Financial Bid)	10. PCR Machine	4. Should have the feature of dynamic ramping (identical hold times) for all the 8 rows of a gradient	Should have the feature of dynamic ramping (identical hold times) for all the 8 rows of a gradient	M/s Eppendorf	Should be "Gradient in 8 rows/Gradient in 12 Column or better"	Should have Gradient in 8 rows/Gradient in 12 Column or better
			5. Should have touch screen interface which can displays	Should have touch screen interface which can	M/s Eppendorf	Should be "Touch Screen/Color Screen or better"	Touch Screen/Color Screen or better

	& 20 (Technical Compliance Report of Financial Bid)		graphics in high resolution for easy programming	displays graphics in high resolution for easy programming			
			7.Should have a maximum ramp rate of 40 – 1000 C or better	7.Should have a maximum ramp rate of 40 – 1000 C or better	M/s Eppendorf	Should be “ Heating rate 30 C and Cooling Rate 20 C or better”	Heating rate 30 C and Cooling Rate 20 C or better